



REV	DESCRIPTION	DRAWN	DATE

ITEM	PART NO.	DESCRIPTION OR SIZE	QTY.
5	COM1L	TUBING, 3/8 O.D. x .049 WALL x 2.00 LG. STN. STL. 304	1
4	COM1L	TUBING, 3/8 O.D. x .049 WALL x .96 LG. STN. STL. 304	1
3	COM1L	TUBING, 3/8 O.D. x .049 WALL x 3.96 LG. STN. STL. 304	1
2	COM1L	TEE, 1.5"x1.5", 3/8 O.D. TUBE, TRUPLY TUBULAR BRAZ-TYPE CAT. NO. 30PT-8, 316 S.S.	1
1	COM1L	WELD FLANGE, 3/8" NOM., .150 NM16.	3

PARTS LIST			
UNLESS OTHERWISE SPECIFIED	ORIGINATOR	R.L. SCHMITT	19-FEB-2002
.XX	ANGLES	W. CYKO	19-FEB-2002
± .01	±	CHECKED	
±	±	APPROVED	
1. BREAK ALL SHARP EDGES TO ROUNDS			
2. DIMENSIONS BASED UPON FULL SCALE DRAWING			
3. DIMENSIONS BASED UPON ANSI Y14.5M-1982 UNLESS OTHERWISE SPECIFIED			
4. ALL DIMENSIONS IN INCHES			
5. MAX ALL MACH. SURFACES			
125			
MATERIAL USED ON MD-384419			
SEE ABOVE PARTS LIST			

 FERMI NATIONAL ACCELERATOR LABORATORY UNITED STATES DEPARTMENT OF ENERGY			
PPD/MECHANICAL DEPT. CDMS MIXTURE PURIFICATION PANEL TUBE & FITTING WELDMENT "B"			
SCALE	DRAWING NUMBER	SHEET	REV
FULL	9213.300-MC-397610	1 OF 1	1
CREATED WITH :		GROUP :	
Ideas8m4		PPD/MECHANICAL DEPARTMENT	



1. VACUUM NOTE: NO LEAK SHALL BE DETECTABLE ON THE MOST SENSITIVE SCALE OF A HELIUM LEAK DETECTOR WITH A MINIMUM SENSITIVITY OF 10⁻⁹ ATM CC/SEC.

NOTES: